



# Certificate of Traceability

## Level I Laboratory Certification

This certifies that LabCorp Central Laboratory, using Bio-Rad Variant II Turbo 2.0 has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the Diabetes Control and Complications Trial Reference method. The comparison was performed with: University of Missouri SRL#9

The system evaluated was:

Instrument: <b>Variant II Turbo 2.0</b>	Calibrator Lot: <b>64463748, 64463749</b>	Column Lot: <b>90072AN</b>
Reagent Lot: <b>64487256, 64473800, 64497659</b>	Calibrator Assigned Values: <b>5.4%, 10.1%</b>	

Date of Certification: May 1, 2023

Certification Expires: May 1, 2024

*David Sork*

NGSP Steering Committee Chair

*Randi Little PhD.*

NGSP Network Coordinator

*Shan Anulf*

SRL director/ supervisor